

# SpermCryo™ All-round

## PRODUCT INFORMATION

### SpermCryo™ All-round

Catalogue no.	
SPCA-0005	5 ml
SPCA-0010	10 ml
SPCA-0020	20 ml

#### Application

SpermCryo™ All-round is a four times concentrated ready-to-use HEPES buffered cryopreservation medium for freezing human semen at ultra low temperatures (-196°C) in liquid nitrogen.

#### Composition

SpermCryo™ All-round is a HEPES buffered freezing medium for human sperm. It contains 26,7% glycerol in order to protect the sperm from damage due to the freezing process and it contains 0,4% HSA.

#### Material not included

- Sperm freezing straws (e.g. CBS high security sperm straws)
- Freezing tank with liquid nitrogen
- LAF Bench (ISO Class 5)
- Sterile pipettes

#### Quality Control

- pH: 7,20 – 7,90  
(Release criteria: 7,20 – 7,60)
- Endotoxin: < 25 EU/ml
- Sterility: sterile SAL 10<sup>-3</sup>
- Sperm Survival Test: ≥ 80% survival after 24h exposure of untreated sperm to the test medium
- Chemical composition
- Use of Ph Eur or USP grade products if available
- Not MEA tested
- Certificate of Analysis available upon request

#### Sterility

SpermCryo™ All-round is sterilized by sterile filtration.

**STERILE A**

#### Precautions and warnings

Always work under hygienic conditions (LAF-bench, ISO Class 5) to avoid possible contamination. Always wear protective clothing when working with specimens. Handle specimens as if capable of transmitting HIV or hepatitis.

Standard measures to prevent infections resulting from the use of medicinal products prepared from human blood or plasma include selection of donors, screening

of individual donations and plasma pools for specific markers of infection and the inclusion of effective manufacturing steps for the inactivation/removal of viruses. Despite this, when medicinal products prepared from human blood or plasma are administered, the possibility of transmitting infective agents cannot be completely excluded. This also applies to unknown or emerging viruses and other pathogens. There are no reports of proven virus transmissions with albumin manufactured to European Pharmacopoeia specifications by established processes.

The above medicinal products do not contain antibiotics.

#### Pre-use checks

Do not use if the seal on the bottle is broken or open when the product is delivered. Do not use if the product shows any signs of microbial contamination or becomes cloudy.

#### Storage Conditions

Store refrigerated (2-8°C). Do not use after expiration date. Do not freeze before use. Keep away from sunlight. After opening the container, do not use the product longer than 24h. Sterile conditions must be maintained and product must be stored at 2° - 8° C. Stable after transport (up to 5 days) at elevated temperatures (≤ 37°C).

#### Technical Support



Gynotec B.V.

Jonckherenhof 7 – 6581 GC Malden – The Netherlands

Phone (+31) 24 3586582 – Fax (+31) 24 3581355

info@gynotec.nl – www.gynotec.com

CE  
0344

Distributor:

- Cut off the end of the straw, place open end inside

# SpermCryo™ All-round

## Instructions for use

### Sperm Preparation

SpermCryo™ All-round can be used in combination with SpermFilter® ready-to-use gradients.

#### Before freezing

It is recommended to concentrate the sperm before freezing in case of very low sperm concentrations. This may increase the sperm quality after thawing and will reduce the number of straws to be frozen.

#### After thawing

In case needed, use sperm preparation techniques after thawing to eliminate dead sperm cells and debris. Dilute the concentrated sperm in a washing medium (e.g. sperm washing medium) to your use.

a container (e.g. test tube) and tap straw against the side of the container in order to allow complete evacuation of the mixture into the container.

Dilute the concentrated sperm in a suitable insemination medium, at least 3 ml per 0.5 ml semen mix thoroughly.

- Centrifuge at 300-350g for 10 minutes.
- Resuspend the pellet in a suitable insemination medium (e.g. Sperm Wash).

### Method

Ensure that all media is well mixed before use.

### Freezing








- Allow the semen to liquefy at room temperature for 30 minutes.
- Add 1 part SpermCryo™ All-round to 4 parts semen.
- Add SpermCryo™ All-round dropwise while gently swirling the sample.

**Caution:** To avoid cold-shock make sure SpermCryo™ All-round is at room temperature.

- Leave the sample for 10 minutes at room temperature for equilibration.
- Suck the mixture into the freezing straws, leaving approximately 1.5 cm of air at the end of the straw. Seal the straws.
- Dry each straw with a linen free cloth and shake in order to move the air bubble to the centre of the straw.
- Freeze the straws vertically just above the liquid nitrogen level for 15 minutes.
- Store the straws in liquid nitrogen.

### Thawing

- Retrieve the required straws from the liquid nitrogen.
- Place the straws in tap water for 5 minutes.

SYMBOL	MEANING
	Catalogue number
	Batch code
	Use by (expiry date)
	Temperature limitations
	Sterile medical device processed using aseptic technique (filtration)
	Consult instructions for use
	CE mark

### Bibliography

- 1 Mahadevan M, Trounson AD. Effect of cryoprotective media and dilution methods on the preservation of human spermatozoa. *Andrologia*, 1983; 15: 355-66.
- 2 Mahadevan M, Trounson AD, Leeton JF. Successful use of human semen cryobanking for in vitro fertilization. *Fertil Steril*, 1983; 15: 355-66.
- 3 Brotherton J. Cryopreservation of human semen. *Archives of Andrology*, 1990; 25: 181-95.

4 Kobayashi T, Kaneko S, Hara I, Park YJ, et al. Concentrating human sperm before cryopreservation. *Andrologia*, 1991; 23: 25-8.

5 Graczykowski JW, Siegel MS. Influence of sperm processing on the fertilizing capacity and recovery of motile sperm from thawed human semen. *Archives of Andrology*, 1991; 26: 155-61.

**VOID**